

(24-6226)

10 September 2024

Page: 1/2

**Committee on Technical Barriers to Trade** 

Original: English

#### NOTIFICATION

The following notification is being circulated in accordance with Article 10.6  $\,$ 

1.	Notifying Member: <u>EGYPT</u>
	If applicable, name of local government involved (Article 3.2 and 7.2):
2.	Agency responsible:
	Egyptian Organization for Standardization and Quality 16 Tadreeb El-Modarrebeen St., Ameriya, Cairo – Egypt E-mail: <u>eos@idsc.net.eg</u> / <u>eos.tbt@eos.org.eg</u> Website: <u>http://www.eos.org.eg</u> Tel.: + (202) 22845528 Fax: + (202) 22845504
	Name and address (including telephone and fax numbers, email and website addresses, if available) of agency or authority designated to handle comments regarding the notification shall be indicated if different from above:
3.	Notified under Article 2.9.2 [X], 2.10.1 [ ], 5.6.2 [ ], 5.7.1 [ ], 3.2 [ ], 7.2 [ ], other:
4.	<b>Products covered (HS or CCCN where applicable, otherwise national tariff heading. ICS numbers may be provided in addition, where applicable):</b> Aids and adaptation for moving (ICS code(s): 11.180.10)
5.	<b>Title, number of pages and language(s) of the notified document:</b> Draft of Egyptian standard for "Walking aids manipulated by both arms — Requirements and test methods — Part 3: Walking tables" (29 page(s), in English)
6.	<b>Description of content:</b> This draft of Egyptian standard specifies requirements and methods of testing the static stability, braking capabilities, static strength and fatigue of walking tables without accessory equipment, unless specified in the particular test procedure. It also gives requirements relating to safety, ergonomics and performance, marking, labelling and information supplied by the manufacturer.
	ISO 11199-3:2005 includes all walking tables with three or more wheels or tips against the walking surface and having arm supports in the shape of a horizontal supporting table or two horizontal forearm supports.
	The requirements and tests are based on everyday usage of walking tables as walking aids, for a maximum user mass as specified by the manufacturer.
	ISO 11199-3:2005 includes walking tables specified for a user mass of not less than 35 kg.
	Worth mentioning is that this Draft standard adopts the technical content of ISO 11199-3:2005 (confirmed in 2022).
7.	<b>Objective and rationale, including the nature of urgent problems where applicable:</b> Safety requirements. ; Quality requirements

#### - 2 -

## 8. Relevant documents:

ISO 11199-3:2005 (confirmed in 2022).

### 9. **Proposed date of adoption:** To be determined

### Proposed date of entry into force: To be determined

**10.** Final date for comments: 60 days from notification

# **11.** Texts available from: National enquiry point [X] or address, telephone and fax numbers and email and website addresses, if available, of other body:

Egyptian Organization for Standardization and Quality Address: 16 Tadreeb El-Modarrebeen St., Ameriya, Cairo- Egypt E-mail: <u>eos@idsc.net.eg</u> / <u>eos.tbt@eos.org.eg</u> Website: <u>http://www.eos.org.eg</u> Tel: + (202) 22845528 Fax: + (202) 22845504