

10 March 2015

(15-1347)

Page: 1/2

Committee on Technical Barriers to Trade

Original: English

NOTIFICATION

The following notification is being circulated in accordance with Article 10.6

<p>1. Notifying Member: <u>EGYPT</u> If applicable, name of local government involved (Article 3.2 and 7.2):</p>
<p>2. Agency responsible: Egyptian Organization for Standardization and Quality 16 Tadreeb El-Modarrebeen St., Ameriya Cairo - Egypt E-mail: moi@idsc.net.eg / tbteos@eos.org.eg Website: http://www.eos.org.eg Tel.: + (202) 22845528 Fax: + (202) 22845504</p> <p>Name and address (including telephone and fax numbers, email and website addresses, if available) of agency or authority designated to handle comments regarding the notification shall be indicated if different from above:</p>
<p>3. Notified under Article 2.9.2 [X], 2.10.1 [], 5.6.2 [], 5.7.1 [], other:</p>
<p>4. Products covered (HS or CCCN where applicable, otherwise national tariff heading. ICS numbers may be provided in addition, where applicable): Infusion equipment for medical use</p>
<p>5. Title, number of pages and language(s) of the notified document: Ministerial Decree No.975/2014 (2 pages,in Arabic) mandating Egyptian Standard ES 4609-5/2014 for "Infusion Equipment for Medical Use - P 5: Burette Infusion Sets for Single Use, Gravity Feed" (11 pages, in Arabic)</p>
<p>6. Description of content: This Ministerial Decree mandates that the producers and importers must comply with ES 4609-5/2014 for "Infusion Equipment for Medical Use – P 5: Burette Infusion Sets for Single Use, Gravity Feed", this Egyptian Standard supersedes its last edition in 2005 and it specifies requirements for types of single-use, gravity feed burette infusion sets of 50ml, 100ml and 150ml nominal capacity for medical use in order to ensure compatibility of use with containers for infusion solutions and intravenous equipment.</p> <p>This standard complies with ISO: 8536-5/2004.</p>
<p>7. Objective and rationale, including the nature of urgent problems where applicable: Test procedures</p>
<p>8. Relevant documents:</p> <ul style="list-style-type: none"> • Ministerial Decree No.975/2014 • ISO: 8536-5/2004 • Infusion Equipment for Medical Use - P 5: Burette Infusion Sets for Single Use, Gravity Feed

9.	Proposed date of adoption: 18 December 2014 Proposed date of entry into force: 18 June 2015
10.	Final date for comments: 60 days from the date of notification
11.	Texts available from: National enquiry point [] or address, telephone and fax numbers and email and website addresses, if available, of other body: Egyptian Organization for Standardization and Quality Address: 16 Tadreeb El-Modarrebeen St., Ameriya Cairo - Egypt Tel: (+202) 22845528 Fax: (+202) 22845504 E-mail: eos@idsc.net.eg / tbt@eos.org.eg Website: http://www.eos.org.eg